



**Catatan Kualifikasi Analisis
(Instrumen KCKT)**

No. : F-AM-04-19

Rev. : 00

Berlaku : 150901

Jenis Pengujian :

Nama Sampel :

Nomor Sampel :

Bobot teoritis (W) :

Instrumen :

Status kalibrasi :

Kolom :

Baku Kerja (BK) :

No. Kode :

Bobot BK, mg :

Faktor BK :

Fp :

1. System Suitability testing

Replikasi	Waktu Retensi (menit)	Area Response	Tailing factor	Plates	Kriteria Penerimaan (sesuai metode)
1					RSD area maks 2,0%
2					
3					
4					
5					
Rata-rata, %					
Rsd, %					

2. Data Pengujian

Replikasi	Bobot Sampel (Wu), mg	Area Response	Kadar, %	Perhitungan
1				Kadar (%) = $\frac{R_u \times W_s \times f \times f_p \times W}{R_s \times W_u \times 450} \times 100\%$
2				
3				
4				
5				
6				
		Rata-rata		
		Rsd, %		
Tgl. Uji :		Analisis :		Paraf :

Kadar teoritis *) =

$$\begin{aligned} \% \text{ Recovery *)} &= \frac{\text{Rata - rata hasil pengujian}}{\text{Kadar teoritis}} \times 100\% \\ &= \dots \% \end{aligned}$$

Keterangan :

*) Diisi oleh evaluator

(Asman IPC & Peng. Produk / MA)