

 INDOFARMA	Catatan Hasil Kalibrasi Internal <i>Internal Calibration Record</i> Automatic Titrator		No. : F-PM-01-09
			Rev. : 00
			Berlaku : 03 September 2014

No. Fasilitas / <i>Facility No.</i>	:	Lokasi / <i>Location</i>	:
No. Kode Kalibrasi / <i>Calibration code</i>	:	Bidang / <i>Department</i>	:
Merek / <i>Brand</i>	:	Suhu ruangan / <i>Room temp.</i>	:
Tipe / <i>Type</i>	:	Kelembaban udara / <i>RH</i>	:
No seri / <i>Serial Number</i>	:	No. Protap Cara Kalibrasi / <i>SOP</i>	:
Kapasitas / <i>Capacity</i>	:	Nama Petugas / <i>Name</i>	:

Kalibrator yang digunakan <i>Calibrator</i>	Tanggal Kalibrasi Kalibrator <i>Cal. date of Calibrator</i>	Tanggal Rekalibrasi Kalibrator <i>Recal. date of Calibrator</i>

1. **Tes ketepatan volume buret / Accuracy of Burette Volume**

[illegible]

Persyaratan/ Requirement :

Keterangan / Note: $F = C \times E$ $G = [B - F]$



Catatan Hasil Kalibrasi Internal
Internal Calibration Record
Automatic Titrator

No. : F-PM-01-00
Rev. : 00
Berlaku : 03 September 2014

2. Test System Performance

Tanggal / Date :
Suhu Ruangan / Room Temp. :
RH :

No.	Bobot sampel/ Sample Weight (mg)	Volume Titrator (ml)	N NaOH
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			
10.			
Rata-rata/ Average	:		
SD	:		
RSD (%)	:		
Persyaratan / Requirement	:		
Kesimpulan/ Conclusion	:		
Paraf Petugas/ Signature	:		
Paraf Asman/ Sign. of Asst. Manager	:		