

 INDOFARMA	Catatan Hasil Kalibrasi Internal <i>Internal Calibration Note</i> Viscosity meter	No. : F-PM-01-32
		Rev. : 00
		Berlaku : 03 September 2014

No. Fasilitas / <i>Facility No.</i>	:	Lokasi / <i>Location</i>	:
No. Kode Kalibrasi / <i>Calibration code</i>	:	Bidang / <i>Department</i>	:
Merek / <i>Brand</i>	:	Suhu ruangan / <i>Room temp.</i>	:
Tipe / <i>Type</i>	:	Kelembaban udara / <i>RH</i>	:
No seri / <i>Serial Number</i>	:	No. Protap Cara Kalibrasi / <i>SOP</i>	:
Kapasitas / <i>Capacity</i>	:	Nama Petugas / <i>Name</i>	:

Kalibrator yang digunakan <i>Calibrator</i>	Tanggal Kalibrasi Kalibrator <i>Cal. date of Calibrator</i>	Tanggal Rekalibrasi Kalibrator <i>Recal. date of Calibrator</i>
1.		

1. Pengukuran Kecepatan/ Calibration of Speed (RPM)

Tanggal/ Date :

Menit ke/ <i>Minute</i>	RPM			
	10	20	50	100
Rata-rata/ <i>Average</i>				
Syarat/ <i>Requirement</i>	9.2-10.2	19.6-20.4	49.0-51.0	98.0-102.0
Kesimpulan/ <i>Conclusion</i>				
Paraf Petugas/ <i>Signature</i>				
Paraf Asman/ <i>Sign. of Asst. Manager</i>				

	Catatan Hasil Kalibrasi Internal <i>Internal Calibration Note</i> Viscosity meter	No. : F-PM-01-32
		Rev. : 00
		Berlaku : 03 September 2014

2. Kalibrasi Viskositas / Calibration of Viscosity

Tanggal/ Date :

Pembanding/ Reference :					
Produsen/ Producer :					
Viskositas/ Viscosity (cps) :			Viskositas/ Viscosity (cps) :		
Suhu pengukuran/ Temp. (°C) :			Suhu pengukuran/ Temp. (°C) :		
No.	t °C	Viscosity (cps)	No.	t °C	Viscosity (cps)
Rata-rata/ Average			Rata-rata/ Average		
Syarat/ Requirement		484.11 - 493.83	935.55 - 954.45		
Kesimpulan/ Conclusion		MS	MS		
Paraf Petugas/ Signature					
Paraf Asman/ Sign. of Asst. Manager					