

	Catatan Hasil Kalibrasi Internal <i>Internal Calibration Note</i> Granulate Flow Tester	No. : F-PM-01-34
		Rev. : 00
		Berlaku/ Date : 3 September 2014

No. Fasilitas / <i>Facility No.</i>	:	Lokasi / <i>Location</i>	:
No. Kode Kalibrasi / <i>Calibration code</i>	:	Bidang / <i>Department</i>	:
Merek / <i>Brand</i>	:	Suhu ruangan / <i>Room temp.</i>	:
Tipe / <i>Type</i>	:	Kelembaban udara / <i>RH</i>	:
No seri / <i>Serial Number</i>	:	No. Protap Cara Kalibrasi / <i>SOP</i>	:
Kapasitas / <i>Capacity</i>	:	Nama Petugas / <i>Name</i>	:

Kalibrator yang digunakan <i>Calibrator</i>	Tanggal Kalibrasi Kalibrator <i>Cal. date of Calibrator</i>	Tanggal Rekalibrasi Kalibrator <i>Recal. date of Calibrator</i>

1. Kalibrasi Neraca / *Calibration of Balance*

Tanggal/Date :

No.	Standard Weight (g)	Actual (g)	Syarat / Requirement (g)	Kesimpulan/ Conclusion	Paraf Petugas/ Signature	Paraf Asman/ Sign. of Asst. Manager
1	50		49.95 – 50.05			
2	100		99.90 – 100.10			
3	150		149.85 – 150.15			

2. Kalibrasi Penunjuk Waktu / *Calibration of Timer*

Tanggal/ Date :

No.	Nominal (detik/second)	Actual (detik/second)	Syarat (detik/second)	Kesimpulan/ Conclusion	Paraf Petugas/ Signature	Paraf Asman/ Sign. of Asst. Manager
1	10		9 – 11			
2	20		19 – 21			
3	30		29 - 31			

3. Kalibrasi Sudut Repose / *Calibration of Angle of Repose*

Tanggal/ Date :

No.	Nominal	Actual	Syarat	Kesimpulan/ Conclusion	Paraf Petugas/ Signature	Paraf Asman/ Sign. of Asst. Manager
1	45°		44° - 46°			
2	45°		44° - 46°			
3	45°		44° - 46°			